

DATA PROOF POINT SALIVA vs. SWAB

SDNA -1000 SALIVA COLLECTION DEVICE



SPECTRUM™

SOLUTIONS



RUCDR Infinite Biologics COVID-19 QPCR Testing

SPECTRUM SDNA -1000 SALIVA COLLECTION DEVICE

Swab vs. Saliva Biosamples



SDNA -1000

STUDY DESIGN



- 50 Subjects with KNOWN COVID-19 status (Positive and Negative)
- Independently verified clinical status
- All in-patients across 3 hospitals
- Swab collection in Universal Transfer Media
- Saliva collection in Spectrum self-collection device
- All samples processed within 72hrs of collection
- Same volume of UTM and saliva input for nucleic acid extraction (300ul)
- Same volume of extraction product (5ul) used for QPCR detection

} Sampled at
same time

EUA FDA & STATE CERTIFICATIONS




EUA#200090
EUA#200091

- EUA submission for both swab and saliva collection
 - Copan/Puritan/Hologic swabs
 - **Spectrum SDNA-1000 saliva collection device**
 - Automated Chemagen Viral DNA/RNA bead-based extraction chemistry
 - ThermoFisher TaqPath COVID-19 detection kit (20ul)
 - Perkin Elmer Coronavirus nucleic acid detection kit qPCR (20ul)
 - Thermo QuantStudio (384), Via7 (384) cyclers
- NJ DOH approval received
- NY DOH approval received
- Testing initiated across hospitals, telehealth, healthcare networks


SALIVA

Internal Control Spike

MS2 - Positive




MS2 - Negative




N Gene

N- Positive




N- Negative




ORF1ab

ORF1 ab Positive




ORF1 ab Negative




S Gene

S - Positive




S-Negative




SWABS

Internal Control Spike

MS2 - Positive




MS2 - Negative




N Gene

N- Positive




N- Negative




ORF1ab

ORF1 ab Positive




ORF1 ab Negative




S Gene

S - Positive







S-Negative







POSITIVE & NEGATIVE SALIVA, POSITIVE SWAB, POSITIVE CONTROL, NEGATIVE CONTROL, NTC

Internal Control Spike



POSITIVE SALIVA & SWAB


Internal Control Spike



DATA SUMMARY

- 100% Positive and Negative agreement for contrived samples
- 100% Positive agreement for 1-2x LoD (4 copies/ml)
- 100% Agreement on all clinical samples
- 100% Agreement on swab vs. saliva results in clinical samples

Sample Test Interpretation				
IC/MS2: JUN	N Gene /VIC (Ct)	ORF1ab Gene/FAM (Ct)	S Gene/ABY (Ct)	Result
<37	Undetermined	Undetermined	Undetermined	Negative
<37	Two or Three targets are <37			Positive
<37	One Target is <37			Re-Test; If re-test is the same, then Inconclusive
Undetermined	Undetermined	Undetermined	Undetermined	Re-Test; If re-test is the same, then Inconclusive
Control Requirements				
	IC/MS2: JUN	N Gene /VIC (Ct)	ORF1ab Gene/FAM (Ct)	S Gene/ABY (Ct)
Negative Control	≤40	Undetermined	Undetermined	Undetermined
Positive Control	Undetermined	Three targets are <37		
NTC	Undetermined	Undetermined	Undetermined	Undetermined



DATA SUMMARY

Contrived Sample Validation

ThermoFisher COVID-19 TaqPath Kit			
Observed Result	Expected Result		
	Positive	Negative	Total
Positive	30	0	30
Negative	0	30	30
Total			60
95% Confidence Interval			
Positive % Agreement for 1-2x LoD			100%
Positive % Agreement for other concentrations			100%
Negative % Agreement			100%

Clinical Sample Validation

ThermoFisher New Coronavirus Nucleic Acid Detection Kit			
Observed Result	Expected Result		
	Positive (NP/Sal)	Negative (NP/Sal)	Total
Positive	30/30	0/0	30/30
Negative	0/0	30/30	30/30
Total			120
95% Confidence Interval			
Positive % Agreement for Clinical Patient NP vs Saliva			100%
Negative % Agreement			100%

Discover your advantage.



SPECTRUMTM
SOLUTIONS